



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994 & ANSI/NCSL Z540-3-2006

LAW CALIBRATION, LLC  
41 Spring Hill Road  
Saco, ME 04072  
Louis A. Waterhouse III Phone: 207 583 2980

CALIBRATION

Valid To: February 29, 2020

Certificate Number: 2398.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,6</sup>:

IV. Chemical

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
pH Meters	(4 to 10) pH	0.003 pH	Buffer solutions
Conductivity Meters	1 µS 10 µS 100 µS 1000 µS 16.67 mS	1 % 1 % 1 % 1 % 1 %	Conductivity solutions

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Calipers <sup>3</sup>	Up to 60 in	360 µin	Caliper master/gage blocks/length standards
Dial Indicators <sup>3</sup>	Up to 1 in	300 µin	Gage blocks
Test Indicators <sup>3</sup>	Up to 1 in	32 µin	Gage blocks

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Height Gages <sup>3</sup>	Up to 40 in	75 µin	Gage blocks and length standards
Micrometers <sup>3</sup>	Up to 12 in	40 µin	Gage blocks and length standards
Rulers & Tape Measures	Up to 60 in	360 µin	Gage blocks
Optical Comparators – X & Y Magnification	Up to 12 in Up to 100x	20 µin 1 µin	Gage blocks
Microscopes – X & Y	Up to 4 in	40 µin	Stage micrometer

## II. Electrical – DC/Low Frequency

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Measure  (1 to 10) mV	(1 to 40) Hz (0.04 to 1) kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	0.03 % + 4 µV 0.02 % + 1.5 µV 0.03 % + 1.5 µV 0.03 % + 1.5 µV 0.5 % + 1.5 µV 4 % + 2.5 µV	Agilent 3458A
100 mV to 1 V	(1 to 40) Hz (0.04 to 1) kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (1 to 2) MHz	0.007 % + 0.5 mV 0.007 % + 0.3 mV 0.014 % + 0.3 mV 0.03 % + 0.3 mV 0.08 % + 0.3 mV 0.3 % + 1.1 mV 1.5 % + 1.1 mV	



Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Measure (cont)			
(10 to 100) V	(1 to 40) Hz (0.04 to 1) kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	0.03 % + 5 mV 0.02 % + 3 mV 0.02 % + 3 mV 0.035 % + 3 mV 0.12 % + 3 mV 0.4 % + 10 mV	Agilent 3458A
(100 to 1000) V	(1 to 40) Hz (0.04 to 1) kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.04 % + 4 mV 0.04 % + 4 mV 0.06 % + 2 mV 0.12 % + 2 mV 0.3 % + 3 mV	
AC Voltage – Generate			
(1 to 33) mV	(10 to 45) Hz (0.045 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.08 % + 6 μV 0.15 % + 6 μV 0.02 % + 6 μV 0.1 % + 6 μV 0.35 % + 12 μV 0.8 % + 50 μV	Fluke 5522A
(33 to 330) mV	(10 to 45) Hz (0.045 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.03 % + 8 μV 0.013 % + 8 μV 0.015 % + 8 μV 0.035 % + 8 μV 0.08 % + 32 μV 0.2 % + 70 μV	
(0.33 to 3.3) V	(10 to 45) Hz (0.045 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.03 % + 50 μV 0.012 % + 25 μV 0.019 % + 50 μV 0.03 % + 50 μV 0.07 % + 130 μV 0.24 % + 600 μV	
(3.3 to 33) V	(10 to 45) Hz (0.045 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.03 % + 650 μV 0.015 % + 200 μV 0.024 % + 600 μV 0.035 % + 600 μV 0.09 % + 1.6 mV	



Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Generate (cont)			
(33 to 330) V	(10 to 45) Hz (0.045 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.019 % + 2 mV 0.02 % + 6 mV 0.025 % + 6 mV 0.03 % + 6 mV 0.2 % + 50 mV	Fluke 5522A
(330 to 1020) V	(45 to 1000) Hz (1 to 5) kHz (5 to 10) kHz	0.03 % + 10 mV 0.025 % + 10 mV 0.03 % + 10 mV	
AC Current – Measure			
(0.01 to 100) µA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz	0.4 % + 0.03 µA 0.15 % + 0.03 µA 0.06 % + 0.03 µA 0.06 % + 0.03 µA	Agilent 3458A
(1 to 100) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz (5 to 20) kHz (50 to 100) kHz	0.4 % + 0.02 µA 0.15 % + 0.02 µA 0.06 % + 0.02 µA 0.03 % + 0.02 µA 0.06 % + 0.02 µA 0.55 % + 0.15 µA	
(0.1 to 1) A	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz (5 to 20) kHz (50 to 100) kHz	0.4 % + 0.2 mA 0.16 % + 0.2 mA 0.08 % + 0.2 mA 0.1 % + 0.2 mA 0.3 % + 0.2 mA 1 % + 0.4 mA	
AC Current – Generate			
(3.3 to 33) mA	(10 to 20) Hz (20 to 45) Hz (0.045 to 1) kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.18 % + 2 µA 0.09 % + 2 µA 0.04 % + 2 µA 0.08 % + 2 µA 0.2 % + 100 µA 0.4 % + 4 µA	Fluke 5522A



Parameter/Range	Frequency	CMC <sup>2, 4, 5</sup> (±)	Comments
AC Current – Generate (cont)			
(33 to 330) mA	(10 to 20) Hz (20 to 45) Hz (0.045 to 1) kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.18 % + 20 µA 0.09 % + 20 µA 0.04 % + 20 µA 0.1 % + 50 µA 0.2 % + 100 µA 0.4 % + 200 µA	Fluke 5522A
(0.33 to 1.1) A	(10 to 45) Hz (0.045 to 1) kHz (1 to 5) kHz (5 to 10) kHz	0.18 % + 100 µA 0.05 % + 100 µA 0.6 % + 1.0 mA 2.5 % + 5.0 mA	
(1.1 to 3) A	(10 to 45) Hz (0.045 to 1) kHz (1 to 5) kHz (5 to 10) kHz	0.18 % + 100 µA 0.05 % + 100 µA 0.6 % + 1.0 mA 2.5 % + 5.0 mA	
(3 to 11) A	(45 to 100) Hz (0.1 to 1) kHz (1 to 5) kHz	0.6 % + 2.0 mA 0.1 % + 2.0 mA 3 % + 2.0 mA	
(11 to 20.5) A	(45 to 100) Hz (0.1 to 1) kHz (1 to 5) kHz	0.12 % + 5.0 mA 0.15 % + 5.0 mA 3 % + 5.0 mA	
20 A to 1 000 A	(45 to 400) Hz	0.12 %	Fluke 50 turn coil
AC Power – Generate PF=1, 60 Hz	10 mW to 336 W 336 W to 3.06 kW 3.06 kW to 20.9 kW	0.14 % 0.14 % 0.12 %	Fluke 5522A

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
DC Voltage – Measure	(0.01 to 100) mV 100 mV to 1 V (1 to 10) V (10 to 100) V (100 to 1000) V	1.1 µV/V + 3 µV 10 µV/V + 0.3 µV 10 µV/V + 0.05 µV 12 µV/V + 30 µV 12 µV/V + 100 µV	Agilent 3458A



Parameter/Equipment	Range	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
DC Voltage – Generate	0.1 $\mu$ V to 330 mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1000) V	23 $\mu$ V/V + 2 $\mu$ V 13 $\mu$ V/V + 4 $\mu$ V 14 $\mu$ V/V + 35 $\mu$ V 21 $\mu$ V/V + 250 $\mu$ V 21 $\mu$ V/V + 1800 $\mu$ V	Fluke 5522A
DC Current – Measure	100 nA 1 $\mu$ A 10 $\mu$ A 100 $\mu$ A 1 mA 10 mA 100 mA 1 A	0.35 nA/A + 0.04 nA 0.25 $\mu$ A/A + 0.4 $\mu$ A 2.5 $\mu$ A/A + 0.05 $\mu$ A 2.5 $\mu$ A/A + 0.5 $\mu$ A 2.5 $\mu$ A/A + 0.5 $\mu$ A 2.5 $\mu$ A/A + 0.1 $\mu$ A 4 $\mu$ A/A + 0.5 $\mu$ A 10 $\mu$ A/A + 1 $\mu$ A	Agilent 3458A
DC Power – Generate	10 mW to 336 W (336 to 3060) 3060 W to 20.9 kW	0.05 % 0.05 % 0.1 %	Fluke 5522A
AC/DC High Voltage	(1 to 6) kV (1 to 40) kV	1 % 0.4 kV	Fluke-80-k6 Fluke-80-k40
AC/DC Current – High Pot Tester	(0.01 to 100) mA	1 %	Fluke-5320A & 5322A decade resistor
Resistance – Measure	(0.01 to 11) $\Omega$ (11 to 33) $\Omega$ (33 to 110) $\Omega$ (110 to 330) $\Omega$ 330 $\Omega$ to 1.1 k $\Omega$ (1.1 to 3.3) k $\Omega$ (3.3 to 11) k $\Omega$ (11 to 33) k $\Omega$ (33 to 110) k $\Omega$ (110 to 330) k $\Omega$ 330 k $\Omega$ to 1.1 M $\Omega$ (1.1 to 3.3) M $\Omega$ (3.3 to 11) M $\Omega$ (11 to 33) M $\Omega$	0.004 % + 0.011 $\Omega$ 0.003 % + 0.017 $\Omega$ 0.0028 % + 0.017 $\Omega$ 0.0028 % + 0.022 $\Omega$ 0.0028 % + 0.022 $\Omega$ 0.0028 % + 0.22 $\Omega$ 0.0028 % + 0.12 $\Omega$ 0.0028 % + 1.2 $\Omega$ 0.0028 % + 1.2 $\Omega$ 0.0032 % + 12 $\Omega$ 0.0032 % + 12 $\Omega$ 0.006 % + 180 $\Omega$ 0.013 % + 300 $\Omega$ 0.025 % + 5 k $\Omega$	Agilent 3458A



Parameter/Equipment	Range	CMC <sup>2,4,5</sup> (±)	Comments
Resistance – Measure (cont)	(33 to 110) MΩ (110 to 330) MΩ (330 to 1100) MΩ	0.05 % + 6 kΩ 0.3 % + 0.2 MΩ 1.5 % + 1 MΩ	Agilent 3458A
Fixed Points	0.01Ω 0.1 Ω 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ	5 μΩ 30 μΩ 50 μΩ 500 μΩ 5 mΩ 50 mΩ 0.5 Ω 5 Ω 50 Ω 1 kΩ	ESI-SR1-series
Insulation Testers	1 kΩ to 10 TΩ	1 %	Fluke-5322/VLC/40
Capacitance – Generate	(220 to 400) pF (0.4 to 1.1) nF (1.1 to 3.3) nF (3.3 to 11) nF (11 to 33) nF (33 to 110) nF (110 to 330) nF (0.33 to 1.1) μF (1.1 to 3.3) μF (3.3 to 11) μF (11 to 33) μF (33 to 110) μF (110 to 330) μF (0.33 to 1.1) mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	0.5 % + 0.2 nF 0.5 % + 0.1 nF 0.5 % + 0.1 nF 0.25 % + 0.2 nF 0.25 % + 0.2 nF 0.25 % + 0.2 nF 0.25 % + 0.2 nF 0.25 % + 0.2 nF 0.25 % + 4 nF 0.25 % + 0.1 nF 0.45 % + 35 nF 0.45 % + 110 nF 0.45 % + 310 nF 0.45 % + 1020 nF 0.45 % + 3020 nF 0.45 % + 11 μF 0.75 % + 50 μF 1.1 % + 120 μF	Fluke 5522A
Fixed Points	0.01μF 0.1μF 1μF	0.01 % 0.01 % 0.01 %	Std. capacitors



Parameter/Equipment	Range	CMC <sup>2, 5, 7</sup> ( $\pm$ )	Comments
Inductance – Generate	100 mH	0.01 %	Std. inductor
Electrical Simulation of Thermocouples <sup>3</sup> – Generate/Measure			
Type B	(600.0 to 1820) °C	0.44 °C	Fluke 5522A, Fluke 744
Type C	(0.0 to 2316) °C	0.84 °C	
Type E	(-250 to 1000) °C	0.50 °C	
Type J	(-210 to 1200) °C	0.27 °C	
Type K	(-200 to 1372) °C	0.40 °C	
Type N	(-200 to 1300) °C	0.40 °C	
Type R	(0.0 to 1767) °C	0.57 °C	
Type S	(0.0 to 1767) °C	0.47 °C	
Type T	(-250 to 400) °C	0.63 °C	
Electrical Simulation of RTDs <sup>3</sup> –			
Pt 395, 100 $\Omega$	(-200 to 800) °C	0.05 °C	Fluke 5522A
Pt 3926, 100 $\Omega$	(-200 to 630) °C	0.05 °C	
Pt 3916, 100 $\Omega$	(-200 to 630) °C	0.25 °C	
Pt 385, 200 $\Omega$	(-200 to 630) °C	0.16 °C	
Pt 385, 500 $\Omega$	(-200 to 630) °C	0.11 °C	
Pt 385, 1000 $\Omega$	(-200 to 630) °C	0.23 °C	
Pt Ni 385, 120 $\Omega$	(-80 to 260) °C	0.14 °C	
Cu 427, 10 $\Omega$	(-100 to 260) °C	0.3 °C	



III. Mechanical

Parameter/Equipment	Range	CMC <sup>2, 5, 7</sup> (±)	Comments
Pressure <sup>3</sup> – Measure and Measuring Equipment	(-15 to 15) psi (0 to 100) psi (0 to 3000) psi (0 to 10 000) psi	0.04 psi 0.01 psi 1.5 psi 5.3 psi	Pressure modules/gauges Fluke # 700G06 Ashcroft # 2074 & 2089 Fluke # 700P31
Mass	(1 to 100) mg (100 to 500) mg (1 to 100) g (100 to 500) g 1 kg 2 kg 5 kg 10 kg 20 kg 25 kg 50 lb	0.008 mg 0.014 mg 0.28 mg 1.4 mg 2.9 mg 5.8 mg 15 mg 29 mg 58 mg 73 mg 0.003 lb	ASTM (Class 2 to Class 7) NIST Class F, OIML (F1, F2, M1, M2, M3)  Comparators Standard weights ASTM – Class 1
Balances & Scales <sup>3</sup>	Up to 100 g 1000 g 10 kg 100 lb 1000 lb 2000 lb 3000 lb	0.15 mg 1.5 mg 15 mg 0.086 lb 0.10 lb 0.50 lb 3.00 lb	Standard weights ASTM – Class 1  Class F
Torque Wrenches	(10 to 100) ozf·in (5 to 50) lbf·in (50 to 500) lbf·in (250 to 750) lbf·ft (750 to 2000) lbf·ft	0.25 % 0.25 % 0.25 % 0.25 % 0.25 %	AWS torque transducers
Torque Transducers	Up to 250 lbf·ft	0.025 %	Standard weights & torque wheels
Force – Measure and Measuring Equipment	Up to 10 lb Up to 100 lb Up to 1000 lb Up to 10 000 lb Up to 100 000 lb Up to 200 000 lb	0.001 lb 0.02 lb 0.5 lb 15 lb 105 lb 250 lb	Standard weights Class F;  Load cells



## V. Thermodynamic

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Temperature –			
Thermometers	(0 -80) °C	0.025 °C	Hart Scientific SPRT 5614
Thermocouples	(-20 to 0) °C	0.05 °C	
Indicators	(0 to 650) °C	0.25 °C	Fluke 9143, 9144, 7526A,1529
Controllers	(-250 to 1800) °C	0.7 °C	Vaisala-HMP75
Relative Humidity –			
Thermohygrometer	(5 to 90) % RH Up to 40 °C	1 % RH 0.25 °C	Fluke 9132,9133 black body
Infrared Thermometer	(-20 to 500) °C	(2.2) °C	

## VI. Time and Frequency

Parameter/Equipment	Range	CMC <sup>2,5,7</sup> (±)	Comments
Frequency – Measure	(1 to 40) Hz 40 Hz to 10 MHz	0.03 % 0.01 %	Agilent 3458
Frequency – Generate	0.01 Hz to 2 MHz	0.25 mHz/Hz + 5 µHz	Fluke 5522A
Stop Watch & Timers	24 Hours	2.5 s	Cole Parmer-94463-30

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

- <sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC's represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- <sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- <sup>4</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.
- <sup>5</sup> In the statement of CMC, percentages are percentage of reading, unless otherwise indicated.
- <sup>6</sup> This scope meets A2LA's P112 *Flexible Scope Policy*.
- <sup>7</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



## *Accredited Laboratory*

A2LA has accredited

**LAW CALIBRATION, LLC**

*Saco, ME*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program, ANSI/NC SLZ540-1-1994 and ANSI/NC SLZ540.3-2006. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAFC communiqué dated April 2017*).



Presented this 6<sup>th</sup> day of June 2018.

A handwritten signature in black ink, written over a horizontal line.

President and CEO  
For the Accreditation Council  
Certificate Number 2398.01  
Valid to February 29, 2020

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*